

Temperature Compensation Chart

Because the temperature of the tested water will effect the result, you should adjust the *dip time* according the following chart:

°C / °F	Dip Time
0 / 32	60
1 / 34	57
2 / 36	54
3 / 37	51
4 / 39	48
5 / 41	46
6 / 43	43
7 / 45	42
8 / 46	40
9 / 48	38
10 / 50	36
11 / 52	34
12 / 54	33
13 / 55	31
14 / 57	29
15 / 59	28
16 / 60	26
17 / 62	24
18 / 64	23
19 / 66	21
20 / 68	20
21 / 70	18

°C / °F	Dip Time
22 / 72	17
23 / 73	16
24 / 75	15
25 / 77	15
26 / 79	14
27 / 81	14
28 / 82	13
29 / 84	13
30 / 86	12
31 / 88	12
32 / 90	12
33 / 91	12
34 / 93	11
35 / 95	11
36 / 97	11
37 / 99	10
38 / 100	10
39 / 102	10
40 / 104	10

Note: Measure the temperature of the water sample within $\pm 2^{\circ}\text{C}$ / $\pm 4^{\circ}\text{F}$